



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** Tully Environmental, Inc

**SDG No.:** Q3760

**Contract:** TULL01

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Copper  | 21.4           | 20.0               | 107           | 65 - 135                  | P | 12/10/2025       | 13:33            | LB138187      |
|           | Iron    | 105            | 100                | 104           | 65 - 135                  | P | 12/10/2025       | 13:33            | LB138187      |
|           | Lead    | 13.7           | 12.0               | 114           | 65 - 135                  | P | 12/10/2025       | 13:33            | LB138187      |