

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q3760</u>
Contract:	<u>TULL01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	21.3	2.0	1065	-18	22	12/10/2025	13:38	LB138187
	Iron	96400	101000	95	85600	116500	12/10/2025	13:38	LB138187
	Lead	8.91			-12	12	12/10/2025	13:38	LB138187
ICSAB01	Copper	475	511	93	434	588	12/10/2025	13:42	LB138187
	Iron	91100	99300	92	84400	114500	12/10/2025	13:42	LB138187
	Lead	59.8	49.0	122	37	61	12/10/2025	13:42	LB138187
ICSA	Copper	23.6	2.0	1180	-18	22	12/10/2025	13:46	LB138187
	Iron	95500	101000	95	85600	116500	12/10/2025	13:46	LB138187
	Lead	46.9			-12	12	12/10/2025	13:46	LB138187
ICSAB	Copper	514	511	101	434	588	12/10/2025	13:50	LB138187
	Iron	90600	99300	91	84400	114500	12/10/2025	13:50	LB138187
	Lead	72.6	49.0	148	37	61	12/10/2025	13:50	LB138187