

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3760</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3771-02</u>	client id: <u>CompMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3771-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	218		85.1		150	89		P
Iron	ug/L	75 - 125	8010		7140		1500	58		P
Lead	ug/L	75 - 125	451		9.00		500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3760</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3771-02</u>	client id: <u>CompMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3771-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	217		85.1		150	88		P
Iron	ug/L	75 - 125	8190		7140		1500	70		P
Lead	ug/L	75 - 125	454		9.00		500	89		P