

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

|                                   |                                 |                     |                 |                                         |                |
|-----------------------------------|---------------------------------|---------------------|-----------------|-----------------------------------------|----------------|
| <b>Client:</b>                    | <u>Tully Environmental, Inc</u> | <b>Level:</b>       | <u>LOW</u>      | <b>SDG No.:</b>                         | <u>Q3760</u>   |
| <b>Contract:</b>                  | <u>TULL01</u>                   |                     |                 | <b>Lab Code:</b>                        | <u>ACE</u>     |
| <b>Matrix:</b>                    | <u>Water</u>                    | <b>Sample ID:</b>   | <u>Q3771-02</u> | <b>Client ID:</b>                       | <u>CompDUP</u> |
| <b>Percent Solids for Sample:</b> | NA                              | <b>Duplicate ID</b> | Q3771-02DUP     | <b>Percent Solids for Spike Sample:</b> | NA             |

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Copper  | ug/L  | 20               | 85.1          |   | 85.2             |   | 0   |      | P |
| Iron    | ug/L  | 20               | 7140          |   | 7050             |   | 1   |      | P |
| Lead    | ug/L  | 20               | 9.00          |   | 9.56             |   | 6   |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

|                                   |                                 |                     |                   |                                         |                |
|-----------------------------------|---------------------------------|---------------------|-------------------|-----------------------------------------|----------------|
| <b>Client:</b>                    | <u>Tully Environmental, Inc</u> | <b>Level:</b>       | <u>LOW</u>        | <b>SDG No.:</b>                         | <u>Q3760</u>   |
| <b>Contract:</b>                  | <u>TULL01</u>                   |                     |                   | <b>Lab Code:</b>                        | <u>ACE</u>     |
| <b>Matrix:</b>                    | <u>Water</u>                    | <b>Sample ID:</b>   | <u>Q3771-02MS</u> | <b>Client ID:</b>                       | <u>CompMSD</u> |
| <b>Percent Solids for Sample:</b> | NA                              | <b>Duplicate ID</b> | Q3771-02MSD       | <b>Percent Solids for Spike Sample:</b> | NA             |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Copper  | ug/L  | 20                  | 218              |   | 217                 |   | 0   |      | P |
| Iron    | ug/L  | 20                  | 8010             |   | 8190                |   | 2   |      | P |
| Lead    | ug/L  | 20                  | 451              |   | 454                 |   | 1   |      | P |