

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170816BS							
Copper	ug/L	150	154		103	85 - 115	P
Iron	ug/L	1500	1470		98	85 - 115	P
Lead	ug/L	500	475		95	85 - 115	P