

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3760</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tully Environmental, Inc</u>	Datafile :	<u>VN088464.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1205WBS01	Benzene	20	20.6	ug/L	103			65	135	
	Toluene	20	19.9	ug/L	100			70	130	
	Ethyl Benzene	20	18.8	ug/L	94			60	140	
	m/p-Xylenes	40	38.2	ug/L	96			87	111	
	o-Xylene	20	18.9	ug/L	95			87	111	