

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3760</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tully Environmental, Inc</u>	Datafile :	<u>VN088465.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1205WBSD01	Benzene	20	21.5	ug/L	108	5		65	135	20
	Toluene	20	21.7	ug/L	109	9		70	130	20
	Ethyl Benzene	20	20.6	ug/L	103	9		60	140	20
	m/p-Xylenes	40	41.4	ug/L	104	8		87	111	20
	o-Xylene	20	20.8	ug/L	104	9		87	111	20