

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q3760
 Lab File ID: VN088463.D Date Analyzed: 12/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	59227	7.79	320011	9.08	286792	11.84
UPPER LIMIT	118454	8.288	640022	9.577	573584	12.341
LOWER LIMIT	29613.5	7.288	160006	8.577	143396	11.341
EPA SAMPLE NO.						
001 Willets Pt Blvd (Dec)	39477	7.79	181676	9.08	165639	11.84
002 35th Ave (Dec)	47619	7.79	225907	9.08	204147	11.84
VN1205WBL01	52610	7.79	265051	9.08	235330	11.84
VN1205WBS01	55054	7.79	271017	9.08	253115	11.84
VN1205WBSD01	58378	7.79	295792	9.08	269874	11.84

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.