

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TULL01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3760</u>
Lab File ID:	<u>VN088463.D</u>	Date Analyzed:	<u>12/05/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>08:56</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001 Willets Pt Blvd (Dec)	0	0.00				
002 35th Ave (Dec)	0	0.00				
VN1205WBL01	0	0.00				
VN1205WBS01	0	0.00				
VN1205WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.