

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

**METALS CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q3763

MATRIX: Water

METHOD: 6010D

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Serial Dilution Summary (if applicable) Submitted.			✓
4. Laboratory Control Sample Summary (if applicable) Submitted.			✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Matrix Spike (MC0KF1MS) analysis met criteria for all compounds except for Silver due to Chemical Interference during Digestion Process. The Matrix Spike Duplicate (MC0KF1MSD) analysis met criteria for all compounds except for Silver due to Chemical Interference during Digestion Process.			
7. Sample Duplicate Analysis Met QC Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Duplicate (MC0KF1MSD) analysis met criteria for all compounds except for Silver due to Chemical Interference during Digestion Process.			
8. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS: The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

---

QA REVIEW

---

Date