



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	12/09/2025	12:24	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	12:24	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	12:24	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	12:24	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	12:24	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	12:24	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	12:24	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	12:24	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	12:24	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	12:24	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	12:24	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	12:24	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	12:24	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	12:24	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	12:24	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	12:24	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	12:24	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	12:24	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	12:24	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	12:24	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	12:24	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	12:24	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	12/09/2025	12:53	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	12:53	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	12:53	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	12:53	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	12:53	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	12:53	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	12:53	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	12:53	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	12:53	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	12:53	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	12:53	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	12:53	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	12:53	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	12:53	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	12:53	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	12:53	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	12:53	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	12:53	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
CCB02	Aluminum	11.3	+/-50	U	100	P	12/09/2025	13:45	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	13:45	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	13:45	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	13:45	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	13:45	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	13:45	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	13:45	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	13:45	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	13:45	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	13:45	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	13:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	13:45	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	13:45	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	13:45	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	13:45	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	13:45	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	13:45	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	12/09/2025	13:45	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	13:45	LB138174
	Thallium	5.85	+/-20	J	40.0	P	12/09/2025	13:45	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	13:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	13:45	LB138174
CCB03	Aluminum	11.3	+/-50	U	100	P	12/09/2025	15:13	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	15:13	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	15:13	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	15:13	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	15:13	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	15:13	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	15:13	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	15:13	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	15:13	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	15:13	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	15:13	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	15:13	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	15:13	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	15:13	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	15:13	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	15:13	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	15:13	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	15:13	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	15:13	LB138174
	Thallium	5.03	+/-20	J	40.0	P	12/09/2025	15:13	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	15:13	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	15:13	LB138174
CCB04	Aluminum	11.3	+/-50	U	100	P	12/09/2025	16:04	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	16:04	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	16:04	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	16:04	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	16:04	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	16:04	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	16:04	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	16:04	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	16:04	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	16:04	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	16:04	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	16:04	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	12/09/2025	16:04	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	16:04	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	16:04	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	16:04	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	16:04	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	16:04	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
CCB05	Aluminum	11.3	+/-50	U	100	P	12/09/2025	16:54	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	16:54	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	16:54	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	16:54	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	16:54	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	16:54	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	16:54	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	16:54	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	16:54	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	16:54	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	16:54	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	16:54	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	16:54	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	16:54	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	16:54	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	16:54	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	16:54	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	16:54	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
CCB06	Aluminum	11.3	+/-50	U	100	P	12/09/2025	17:45	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	17:45	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	17:45	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	17:45	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	17:45	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	17:45	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	17:45	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	17:45	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	17:45	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	17:45	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	17:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	17:45	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	17:45	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	17:45	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	17:45	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	17:45	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	17:45	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	17:45	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	17:45	LB138174
	Thallium	5.28	+/-20	J	40.0	P	12/09/2025	17:45	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	17:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	17:45	LB138174
CCB07	Aluminum	11.3	+/-50	U	100	P	12/09/2025	18:36	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	18:36	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	18:36	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	18:36	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	18:36	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	18:36	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	18:36	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	18:36	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	18:36	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	18:36	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	18:36	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	18:36	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	18:36	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	18:36	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	18:36	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	18:36	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	18:36	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	18:36	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
CCB08	Aluminum	11.3	+/-50	U	100	P	12/09/2025	19:34	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	19:34	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	19:34	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	19:34	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	19:34	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	19:34	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	19:34	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	19:34	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	19:34	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	19:34	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	19:34	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	19:34	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	19:34	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	19:34	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	19:34	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	19:34	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	19:34	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	19:34	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	19:34	LB138174
	Thallium	5.64	+/-20	J	40.0	P	12/09/2025	19:34	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	19:34	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	19:34	LB138174
CCB09	Aluminum	11.3	+/-50	U	100	P	12/09/2025	20:41	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	20:41	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	20:41	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	20:41	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	20:41	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	20:41	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	20:41	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	20:41	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	20:41	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	20:41	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	20:41	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	20:41	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	20:41	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	20:41	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	20:41	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	20:41	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	20:41	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	20:41	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	20:41	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	7.51	+/-20	J	40.0	P	12/09/2025	20:41	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	20:41	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	20:41	LB138174
CCB10	Aluminum	11.3	+/-50	U	100	P	12/09/2025	21:31	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	21:31	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	21:31	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	21:31	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	21:31	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	21:31	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	21:31	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	21:31	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	21:31	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	21:31	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	21:31	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	21:31	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	21:31	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	21:31	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	21:31	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	21:31	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	21:31	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	21:31	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	21:31	LB138174
	Thallium	8.81	+/-20	J	40.0	P	12/09/2025	21:31	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	21:31	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	21:31	LB138174
	Aluminum	11.3	+/-50	U	100	P	12/09/2025	22:22	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	22:22	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	22:22	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	22:22	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	22:22	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	22:22	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	22:22	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	22:22	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	22:22	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	22:22	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	22:22	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	22:22	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	22:22	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	22:22	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	22:22	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	22:22	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	22:22	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	22:22	LB138174
	Thallium	4.38	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
CCB12	Aluminum	11.3	+/-50	U	100	P	12/09/2025	23:25	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	23:25	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	23:25	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	23:25	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	23:25	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	23:25	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	23:25	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	23:25	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	23:25	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	23:25	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	23:25	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	23:25	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	23:25	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	23:25	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	23:25	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	23:25	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	23:25	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	23:25	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	23:25	LB138174
	Thallium	5.46	+/-20	J	40.0	P	12/09/2025	23:25	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	23:25	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	23:25	LB138174
	Aluminum	11.3	+/-50	U	100	P	12/09/2025	23:45	LB138174
	Antimony	6.76	+/-25	U	50.0	P	12/09/2025	23:45	LB138174
	Arsenic	5.12	+/-10	U	20.0	P	12/09/2025	23:45	LB138174
	Barium	14.6	+/-50	U	100	P	12/09/2025	23:45	LB138174
	Beryllium	0.56	+/-3	U	6.00	P	12/09/2025	23:45	LB138174
	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	23:45	LB138174
	Calcium	234	+/-1000	U	2000	P	12/09/2025	23:45	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	23:45	LB138174
	Cobalt	2.26	+/-15	U	30.0	P	12/09/2025	23:45	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3763

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB13	Copper	4.60	+/-10	U	20.0	P	12/09/2025	23:45	LB138174
	Iron	23.4	+/-50	U	100	P	12/09/2025	23:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	23:45	LB138174
	Magnesium	244	+/-1000	U	2000	P	12/09/2025	23:45	LB138174
	Manganese	5.94	+/-10	U	20.0	P	12/09/2025	23:45	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	23:45	LB138174
	Potassium	918	+/-1000	U	2000	P	12/09/2025	23:45	LB138174
	Selenium	9.64	+/-10	U	20.0	P	12/09/2025	23:45	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	23:45	LB138174
	Sodium	868	+/-1000	U	2000	P	12/09/2025	23:45	LB138174
	Thallium	10.7	+/-20	J	40.0	P	12/09/2025	23:45	LB138174
	Vanadium	6.26	+/-20	U	40.0	P	12/09/2025	23:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	23:45	LB138174