

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

MC0KE4L

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **Lb No.:** lb138174 **Lab Sample ID :** Q3764-10L **SDG No.:** Q3764
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
		C		C			
Aluminum	43.8	J	38.8	J	11		P
Antimony	18.7	J	20.4	J	9		P
Arsenic	10.0	U	50.0	U			P
Barium	18.6	J	250	U	100.0		P
Beryllium	3.00	U	15.0	U			P
Cadmium	10.2		10.1	J	1		P
Calcium	13200		13400		2		P
Chromium	5.00	U	25.0	U			P
Cobalt	15.0	U	75.0	U			P
Copper	5.46	J	50.0	U	100.0		P
Iron	249		258		4		P
Lead	209		208		0		P
Magnesium	5970		5970		0		P
Manganese	17.0		18.2	J	7		P
Nickel	3.85	J	100	U	100.0		P
Potassium	2350		2500	J	6		P
Selenium	10.0	U	50.0	U			P
Silver	5.00	U	25.0	U			P
Sodium	925	J	5000	U	100.0		P
Thallium	2.26	J	12.3	J	442		P
Vanadium	20.0	U	100	U			P
Zinc	19.2	J	100	U	100.0		P