

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3765

Contract: TETR16

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170841BS							
Aluminum	ug/L	1000	1040		104	86 - 115	P
Antimony	ug/L	400	418		104	88 - 113	P
Arsenic	ug/L	400	412		103	87 - 113	P
Barium	ug/L	100	101		101	88 - 113	P
Beryllium	ug/L	100	107		107	89 - 112	P
Cadmium	ug/L	100	102		102	88 - 113	P
Calcium	ug/L	500	528	J	106	87 - 113	P
Chromium	ug/L	200	201		100	90 - 113	P
Cobalt	ug/L	100	101		101	89 - 114	P
Copper	ug/L	150	160		107	86 - 114	P
Iron	ug/L	1500	1460		97	87 - 115	P
Lead	ug/L	500	501		100	86 - 113	P
Magnesium	ug/L	1000	1050		105	85 - 113	P
Manganese	ug/L	100	103		103	90 - 114	P
Nickel	ug/L	250	256		102	88 - 113	P
Potassium	ug/L	5000	4860		97	86 - 114	P
Selenium	ug/L	1000	1070		107	83 - 114	P
Silver	ug/L	37.5	37.7		100	84 - 115	P
Sodium	ug/L	1500	1460		97	87 - 115	P
Thallium	ug/L	1000	1040		104	85 - 114	P
Vanadium	ug/L	150	155		103	90 - 111	P
Zinc	ug/L	100	109		109	87 - 115	P