

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

MC0KH7L

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **Lb No.:** lb138174 **Lab Sample ID :** Q3766-09L **SDG No.:** Q3766
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
		C		C			
Aluminum	10.8	J	250	U	100.0		P
Antimony	10.3	J	125	U	100.0		P
Arsenic	10.0	U	50.0	U			P
Barium	16.4	J	250	U	100.0		P
Beryllium	3.00	U	15.0	U			P
Cadmium	268		280		4		P
Calcium	6310		6460		2		P
Chromium	5.00	U	25.0	U			P
Cobalt	15.0	U	75.0	U			P
Copper	17.1		18.0	J	5		P
Iron	188		169	J	10		P
Lead	570		586		3		P
Magnesium	1320		1330	J	1		P
Manganese	50.6		48.7	J	4		P
Nickel	17.5	J	18.1	J	4		P
Potassium	1650		5000	U	100.0		P
Selenium	11.1		50.0	U	100.0		P
Silver	5.00	U	25.0	U			P
Sodium	1370		5000	U	100.0		P
Thallium	3.95	J	21.6	J	448		P
Vanadium	20.0	U	100	U			P
Zinc	391		399		2		P