

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3771

Project: Monthly 2025

Run No.: LB138103

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138103BS							
BOD5	mg/L	198	217		109	1	84.6-115.4	12/04/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3771

Project: Monthly 2025

Run No.: LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3771

Project: Monthly 2025

Run No.: LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BS							
TPH	mg/L	20.0	18.0		90	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3771

Project: Monthly 2025

Run No.: LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BSD							
TPH	mg/L	20.0	18.3		92	1	78-114	12/08/2025