

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Mercury	3.95	4.0	99	95 - 105	CV	12/16/2025	11:31	LB138234

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV20	Mercury	5.07	5.0	101	90 - 110	CV	12/16/2025	11:36	LB138234
CCV21	Mercury	5.21	5.0	104	90 - 110	CV	12/16/2025	12:29	LB138234
CCV22	Mercury	4.69	5.0	94	90 - 110	CV	12/16/2025	12:43	LB138234

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	2010	2000	100	95 - 105	P	12/10/2025	13:06	LB138187
	Copper	986	1000	99	95 - 105	P	12/10/2025	13:06	LB138187
	Lead	4030	4000	101	95 - 105	P	12/10/2025	13:06	LB138187
	Nickel	2010	2000	101	95 - 105	P	12/10/2025	13:06	LB138187
	Zinc	1970	2000	99	95 - 105	P	12/10/2025	13:06	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.12	6.0	102	80 - 120	P	12/10/2025	13:16	LB138187
	Copper	21.5	20.0	107	80 - 120	P	12/10/2025	13:16	LB138187
	Lead	12.2	12.0	102	80 - 120	P	12/10/2025	13:16	LB138187
	Nickel	40.8	40.0	102	80 - 120	P	12/10/2025	13:16	LB138187
	Zinc	42.4	40.0	106	80 - 120	P	12/10/2025	13:16	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2470	2500	99	90 - 110	P	12/10/2025	13:55	LB138187
	Copper	1250	1250	100	90 - 110	P	12/10/2025	13:55	LB138187
	Lead	4970	5000	99	90 - 110	P	12/10/2025	13:55	LB138187
	Nickel	2480	2500	99	90 - 110	P	12/10/2025	13:55	LB138187
	Zinc	2530	2500	101	90 - 110	P	12/10/2025	13:55	LB138187
CCV02	Cadmium	2420	2500	97	90 - 110	P	12/10/2025	14:45	LB138187
	Copper	1250	1250	100	90 - 110	P	12/10/2025	14:45	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	14:45	LB138187
	Nickel	2450	2500	98	90 - 110	P	12/10/2025	14:45	LB138187
	Zinc	2510	2500	100	90 - 110	P	12/10/2025	14:45	LB138187
CCV03	Cadmium	2450	2500	98	90 - 110	P	12/10/2025	15:46	LB138187
	Copper	1260	1250	101	90 - 110	P	12/10/2025	15:46	LB138187
	Lead	4930	5000	99	90 - 110	P	12/10/2025	15:46	LB138187
	Nickel	2460	2500	98	90 - 110	P	12/10/2025	15:46	LB138187
	Zinc	2500	2500	100	90 - 110	P	12/10/2025	15:46	LB138187
CCV04	Cadmium	2420	2500	97	90 - 110	P	12/10/2025	16:42	LB138187
	Copper	1240	1250	99	90 - 110	P	12/10/2025	16:42	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	16:42	LB138187
	Nickel	2440	2500	98	90 - 110	P	12/10/2025	16:42	LB138187
	Zinc	2490	2500	100	90 - 110	P	12/10/2025	16:42	LB138187
CCV05	Cadmium	2410	2500	96	90 - 110	P	12/10/2025	17:49	LB138187
	Copper	1240	1250	99	90 - 110	P	12/10/2025	17:49	LB138187
	Lead	4860	5000	97	90 - 110	P	12/10/2025	17:49	LB138187
	Nickel	2430	2500	97	90 - 110	P	12/10/2025	17:49	LB138187
	Zinc	2390	2500	96	90 - 110	P	12/10/2025	17:49	LB138187
CCV06	Cadmium	2380	2500	95	90 - 110	P	12/10/2025	18:41	LB138187
	Copper	1220	1250	98	90 - 110	P	12/10/2025	18:41	LB138187
	Lead	4820	5000	96	90 - 110	P	12/10/2025	18:41	LB138187
	Nickel	2410	2500	96	90 - 110	P	12/10/2025	18:41	LB138187
	Zinc	2360	2500	94	90 - 110	P	12/10/2025	18:41	LB138187
CCV07	Cadmium	2370	2500	95	90 - 110	P	12/10/2025	19:33	LB138187
	Copper	1220	1250	98	90 - 110	P	12/10/2025	19:33	LB138187
	Lead	4790	5000	96	90 - 110	P	12/10/2025	19:33	LB138187
	Nickel	2390	2500	96	90 - 110	P	12/10/2025	19:33	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2380	2500	95	90 - 110	P	12/10/2025	19:33	LB138187
CCV08	Cadmium	2400	2500	96	90 - 110	P	12/10/2025	20:21	LB138187
	Copper	1240	1250	99	90 - 110	P	12/10/2025	20:21	LB138187
	Lead	4870	5000	98	90 - 110	P	12/10/2025	20:21	LB138187
	Nickel	2430	2500	97	90 - 110	P	12/10/2025	20:21	LB138187
	Zinc	2410	2500	96	90 - 110	P	12/10/2025	20:21	LB138187