



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.99	6.0	100	65 - 135	P	12/10/2025	13:33	LB138187
	Copper	21.4	20.0	107	65 - 135	P	12/10/2025	13:33	LB138187
	Lead	13.7	12.0	114	65 - 135	P	12/10/2025	13:33	LB138187
	Nickel	40.3	40.0	101	65 - 135	P	12/10/2025	13:33	LB138187
	Zinc	41.3	40.0	103	65 - 135	P	12/10/2025	13:33	LB138187
CRA	Mercury	0.045	0.05	89	60 - 140	CV	12/16/2025	11:40	LB138234