



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Mercury	0.027	+/-0.2	U	0.20	CV	12/16/2025	11:33	LB138234

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB20	Mercury	0.027	+/-0.2	U	0.20	CV	12/16/2025	11:38	LB138234
CCB21	Mercury	0.027	+/-0.2	U	0.20	CV	12/16/2025	12:32	LB138234
CCB22	Mercury	0.027	+/-0.2	U	0.20	CV	12/16/2025	12:45	LB138234

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	13:21	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	13:21	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:21	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	13:21	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	13:21	LB138187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	13:59	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	13:59	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:59	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	13:59	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	13:59	LB138187
CCB02	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	14:49	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	14:49	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	14:49	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	14:49	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	14:49	LB138187
CCB03	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	15:50	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	15:50	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	15:50	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	15:50	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	15:50	LB138187
CCB04	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	16:46	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	16:46	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	16:46	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	16:46	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	16:46	LB138187
CCB05	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	17:53	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	17:53	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	17:53	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	17:53	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	17:53	LB138187
CCB06	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	18:45	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	18:45	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	18:45	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	18:45	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	18:45	LB138187
CCB07	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	19:37	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	19:37	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	19:37	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	19:37	LB138187
	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	19:37	LB138187
CCB08	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	20:25	LB138187
	Copper	4.60	+/-10	U	20.0	P	12/10/2025	20:25	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	20:25	LB138187
	Nickel	3.06	+/-20	U	40.0	P	12/10/2025	20:25	LB138187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	16.7	+/-20	U	40.0	P	12/10/2025	20:25	LB138187