

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Aramark Uniforms</u>	SDG No.: <u>Q3771</u>
Contract: <u>ARAM01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	2.15	1.0	215	-5	7	12/10/2025	13:38	LB138187
	Copper	21.3	2.0	1065	-18	22	12/10/2025	13:38	LB138187
	Lead	8.91			-12	12	12/10/2025	13:38	LB138187
	Nickel	6.33	2.0	316	-38	42	12/10/2025	13:38	LB138187
	Zinc	5.91			-40	40	12/10/2025	13:38	LB138187
ICSAB01	Cadmium	941	972	97	826	1120	12/10/2025	13:42	LB138187
	Copper	475	511	93	434	588	12/10/2025	13:42	LB138187
	Lead	59.8	49.0	122	37	61	12/10/2025	13:42	LB138187
	Nickel	945	954	99	810	1100	12/10/2025	13:42	LB138187
	Zinc	988	952	104	809	1095	12/10/2025	13:42	LB138187
ICSA	Cadmium	0.43	1.0	43	-5	7	12/10/2025	13:46	LB138187
	Copper	23.6	2.0	1180	-18	22	12/10/2025	13:46	LB138187
	Lead	46.9			-12	12	12/10/2025	13:46	LB138187
	Nickel	16.7	2.0	835	-38	42	12/10/2025	13:46	LB138187
	Zinc	-11.5			-40	40	12/10/2025	13:46	LB138187
ICSAB	Cadmium	932	972	96	826	1120	12/10/2025	13:50	LB138187
	Copper	514	511	101	434	588	12/10/2025	13:50	LB138187
	Lead	72.6	49.0	148	37	61	12/10/2025	13:50	LB138187
	Nickel	946	954	99	810	1100	12/10/2025	13:50	LB138187
	Zinc	965	952	101	809	1095	12/10/2025	13:50	LB138187