

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3771</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3771-02</u>	Client ID:	<u>CompDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3771-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	0.81	J	0.81	J	0		P
Copper	ug/L	20	85.1		85.2		0		P
Lead	ug/L	20	9.00		9.56		6		P
Nickel	ug/L	20	24.9		25.1		1		P
Zinc	ug/L	20	504		499		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3771</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3771-02MS</u>	Client ID:	<u>CompMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3771-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	91.4		92.3		1		P
Copper	ug/L	20	218		217		0		P
Lead	ug/L	20	451		454		1		P
Nickel	ug/L	20	254		260		2		P
Zinc	ug/L	20	577		563		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3771</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3816-02</u>	Client ID:	<u>EFFLUENT-METALSDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3816-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.028	J	200.0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3771</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3816-02MS</u>	Client ID:	<u>EFFLUENT-METALSMDS</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3816-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.32		3.30		1		CV