

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170816BS							
Cadmium	ug/L	100	95.5		96	85 - 115	P
Copper	ug/L	150	154		103	85 - 115	P
Lead	ug/L	500	475		95	85 - 115	P
Nickel	ug/L	250	245		98	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q3771

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170963BS Mercury	ug/L	4.0	4.23		106	85 - 115	CV