

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138103

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138103BS							
BOD5	mg/L	198	217		109	1	84.6-115.4	12/04/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138129

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138129BS							
Orthophosphate as P	mg/L	0.5	0.50		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138150BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138125

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170830BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3772

Project: Pre Treatment Plant 2025

Run No.: LB138132

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170844BS							
Phosphorus, Total	mg/L	0.50	0.52		104	1	90-110	12/05/2025