

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q3773
Project:	Linden Generating Station - 402	Sample ID:	Q3760-02
Client ID:	002 35th Ave (Dec)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	131		133		1	1.51		12/04/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q3773
Project:	Linden Generating Station - 402	Sample ID:	Q3773-01
Client ID:	TOWER#1DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	2040		2040		1	0.1		12/03/2025