

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3773

Project: Linden Generating Station - 402

Run No.: LB138101

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138101BS							
TDS	mg/L	100	100		100	1	90-110	12/03/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3773

Project: Linden Generating Station - 402

Run No.: LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025