

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3774

Project: Linden Generating Station - LRSA

RunNo.: LB138124

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	48.150	50	96	95-105	12/05/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	12/05/2025
Sample ID: CCV3 COD	mg/L	49.161	50	98	95-105	12/05/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3774

Project: Linden Generating Station - LRSA

RunNo.: LB138124

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------