

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3774

Project: Linden Generating Station - LRSA

RunNo.: LB138124

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	09/08/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	12/05/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	12/05/2025
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	1.50	10	12/05/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3774

Project: Linden Generating Station - LRSA

RunNo.: LB138124

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------