

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q3776

Project: CTO WA194808 Dahlgren VA - DOD VELAP

RunNo.: LB138146

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	12/08/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	12/08/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	12/08/2025

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q3776

Project: CTO WA194808 Dahlgren VA - DOD VELAP

RunNo.: LB138163

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.1	81	101	78-84	12/09/2025