

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026250.D
 Acq On : 08 Dec 2025 13:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
2	1,4-Dioxane	0.506	0.529	-4.5	68	0.00
3	Pyridine	1.453	1.444	0.6	65	0.00
4	n-Nitrosodimethylamine	0.542	0.529	2.4	64	0.00
5 S	2-Fluorophenol	1.244	1.241	0.2	63	0.00
6	Aniline	2.061	1.917	7.0	64	0.00
7 S	Phenol-d6	1.570	1.512	3.7	63	0.00
8	2-Chlorophenol	1.414	1.388	1.8	62	0.00
9	Benzaldehyde	0.847	1.079	-27.4#	67	0.00
10 C	Phenol	1.689	1.618	4.2	63	0.00
11	bis(2-Chloroethyl)ether	1.324	1.250	5.6	62	0.00
12	1,3-Dichlorobenzene	1.530	1.585	-3.6	65	0.00
13 C	1,4-Dichlorobenzene	1.541	1.585	-2.9	64	0.00
14	1,2-Dichlorobenzene	1.480	1.511	-2.1	64	0.00
15	Benzyl Alcohol	1.142	1.100	3.7	65	0.00
16	2,2'-oxybis(1-Chloropropane	1.848	1.716	7.1	63	0.00
17	2-Methylphenol	1.146	1.106	3.5	64	0.00
18	Hexachloroethane	0.560	0.581	-3.7	65	0.00
19 P	n-Nitroso-di-n-propylamine	0.968	0.920	5.0	65	0.00
20	3+4-Methylphenols	1.583	1.485	6.2	64	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
22	Acetophenone	0.507	0.525	-3.6	64	0.00
23 S	Nitrobenzene-d5	0.351	0.367	-4.6	64	0.00
24	Nitrobenzene	0.355	0.369	-3.9	64	0.00
25	Isophorone	0.689	0.703	-2.0	66	0.00
26 C	2-Nitrophenol	0.180	0.192	-6.7	64	0.00
27	2,4-Dimethylphenol	0.318	0.309	2.8	68	0.00
28	bis(2-Chloroethoxy)methane	0.433	0.430	0.7	64	0.00
29 C	2,4-Dichlorophenol	0.324	0.342	-5.6	64	0.00
30	1,2,4-Trichlorobenzene	0.333	0.362	-8.7	64	0.00
31	Naphthalene	1.081	1.123	-3.9	64	0.00
32	Benzoic acid	0.255	0.268	-5.1	101	0.02
33	4-Chloroaniline	0.453	0.459	-1.3	66	0.00
34 C	Hexachlorobutadiene	0.217	0.241	-11.1	64	0.00
35	Caprolactam	0.114	0.114	0.0	66	-0.01
36 C	4-Chloro-3-methylphenol	0.341	0.349	-2.3	66	0.00
37	2-Methylnaphthalene	0.762	0.789	-3.5	65	0.00
38	1-Methylnaphthalene	0.744	0.778	-4.6	65	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	63	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.611	0.653	-6.9	65	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.408	-7.1	84	0.00
42 S	2,4,6-Tribromophenol	0.260	0.288	-10.8	68	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.439	-5.8	66	0.00
44	2,4,5-Trichlorophenol	0.462	0.487	-5.4	67	0.00
45 S	2-Fluorobiphenyl	1.376	1.492	-8.4	67	0.00
46	1,1'-Biphenyl	1.510	1.539	-1.9	64	0.00
47	2-Chloronaphthalene	1.190	1.242	-4.4	65	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026250.D
 Acq On : 08 Dec 2025 13:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.326	0.334	-2.5	67	0.01
49	Acenaphthylene	1.953	2.058	-5.4	67	0.00
50	Dimethylphthalate	1.490	1.554	-4.3	67	0.00
51	2,6-Dinitrotoluene	0.295	0.312	-5.8	65	-0.01
52 C	Acenaphthene	1.142	1.164	-1.9	66	0.00
53	3-Nitroaniline	0.344	0.340	1.2	65	-0.01
54 P	2,4-Dinitrophenol	0.157	0.163	-3.8	185#	0.00
55	Dibenzofuran	1.776	1.884	-6.1	67	0.00
56 P	4-Nitrophenol	0.307	0.278	9.4	66	0.00
57	2,4-Dinitrotoluene	0.438	0.469	-7.1	68	0.00
58	Fluorene	1.405	1.512	-7.6	68	-0.01
59	2,3,4,6-Tetrachlorophenol	0.391	0.419	-7.2	68	-0.02
60	Diethylphthalate	1.495	1.567	-4.8	68	0.00
61	4-Chlorophenyl-phenylether	0.700	0.759	-8.4	67	-0.01
62	4-Nitroaniline	0.357	0.357	0.0	68	-0.01
63	Azobenzene	1.257	1.279	-1.8	67	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.125	-5.9	100	0.00
66 c	n-Nitrosodiphenylamine	0.618	0.650	-5.2	68	-0.01
67	4-Bromophenyl-phenylether	0.222	0.238	-7.2	67	-0.02
68	Hexachlorobenzene	0.258	0.284	-10.1	68	-0.01
69	Atrazine	0.233	0.239	-2.6	67	0.00
70 C	Pentachlorophenol	0.179	0.184	-2.8	67	0.00
71	Phenanthrene	1.127	1.191	-5.7	69	0.00
72	Anthracene	1.148	1.203	-4.8	67	0.00
73	Carbazole	1.087	1.105	-1.7	67	0.00
74	Di-n-butylphthalate	1.284	1.383	-7.7	69	-0.02
75 C	Fluoranthene	1.372	1.447	-5.5	67	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
77	Benzidine	0.628	0.580	7.6	60	-0.01
78	Pyrene	1.303	1.394	-7.0	65	-0.01
79 S	Terphenyl-d14	0.979	1.110	-13.4	68	0.00
80	Butylbenzylphthalate	0.572	0.605	-5.8	64	0.00
81	Benzo(a)anthracene	1.372	1.425	-3.9	62	0.00
82	3,3'-Dichlorobenzidine	0.499	0.511	-2.4	61	-0.01
83	Chrysene	1.299	1.355	-4.3	62	0.00
84	Bis(2-ethylhexyl)phthalate	0.860	0.907	-5.5	65	0.00
85 c	Di-n-octyl phthalate	1.517	1.601	-5.5	63	0.00
86 I	Perylene-d12	1.000	1.000	0.0	58	0.00
87	Indeno(1,2,3-cd)pyrene	1.512	1.568	-3.7	57	-0.02
88	Benzo(b)fluoranthene	1.215	1.224	-0.7	58	0.00
89	Benzo(k)fluoranthene	1.216	1.267	-4.2	60	-0.01
90 C	Benzo(a)pyrene	1.153	1.195	-3.6	59	-0.01
91	Dibenzo(a,h)anthracene	1.236	1.284	-3.9	57	-0.01
92	Benzo(g,h,i)perylene	1.232	1.272	-3.2	56	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
Data File : BP026250.D
Acq On : 08 Dec 2025 13:44
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0