

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026250.D
 Acq On : 08 Dec 2025 13:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	63	0.00
2	1,4-Dioxane	40.000	41.793	-4.5	68	0.00
3	Pyridine	40.000	39.761	0.6	65	0.00
4	n-Nitrosodimethylamine	40.000	39.052	2.4	64	0.00
5 S	2-Fluorophenol	80.000	79.788	0.3	63	0.00
6	Aniline	40.000	37.209	7.0	64	0.00
7 S	Phenol-d6	80.000	77.053	3.7	63	0.00
8	2-Chlorophenol	40.000	39.263	1.8	62	0.00
9	Benzaldehyde	40.000	50.983	-27.5#	67	0.00
10 C	Phenol	40.000	38.321	4.2	63	0.00
11	bis(2-Chloroethyl)ether	40.000	37.766	5.6	62	0.00
12	1,3-Dichlorobenzene	40.000	41.444	-3.6	65	0.00
13 C	1,4-Dichlorobenzene	40.000	41.126	-2.8	64	0.00
14	1,2-Dichlorobenzene	40.000	40.846	-2.1	64	0.00
15	Benzyl Alcohol	40.000	38.531	3.7	65	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.133	7.2	63	0.00
17	2-Methylphenol	40.000	38.597	3.5	64	0.00
18	Hexachloroethane	40.000	41.531	-3.8	65	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.027	4.9	65	0.00
20	3+4-Methylphenols	40.000	37.510	6.2	64	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	61	0.00
22	Acetophenone	40.000	41.358	-3.4	64	0.00
23 S	Nitrobenzene-d5	80.000	83.482	-4.4	64	0.00
24	Nitrobenzene	40.000	41.550	-3.9	64	0.00
25	Isophorone	40.000	40.799	-2.0	66	0.00
26 C	2-Nitrophenol	40.000	42.799	-7.0	64	0.00
27	2,4-Dimethylphenol	40.000	38.855	2.9	68	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.744	0.6	64	0.00
29 C	2,4-Dichlorophenol	40.000	42.296	-5.7	64	0.00
30	1,2,4-Trichlorobenzene	40.000	43.539	-8.8	64	0.00
31	Naphthalene	40.000	41.530	-3.8	64	0.00
32	Benzoic acid	40.000	42.171	-5.4	101	0.02
33	4-Chloroaniline	40.000	40.521	-1.3	66	0.00
34 C	Hexachlorobutadiene	40.000	44.415	-11.0	64	0.00
35	Caprolactam	40.000	39.854	0.4	66	-0.01
36 C	4-Chloro-3-methylphenol	40.000	41.020	-2.6	66	0.00
37	2-Methylnaphthalene	40.000	41.430	-3.6	65	0.00
38	1-Methylnaphthalene	40.000	41.793	-4.5	65	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	63	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	42.787	-7.0	65	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.851	-7.1	84	0.00
42 S	2,4,6-Tribromophenol	80.000	88.525	-10.7	68	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.313	-5.8	66	0.00
44	2,4,5-Trichlorophenol	40.000	42.128	-5.3	67	0.00
45 S	2-Fluorobiphenyl	80.000	86.744	-8.4	67	0.00
46	1,1'-Biphenyl	40.000	40.773	-1.9	64	0.00
47	2-Chloronaphthalene	40.000	41.753	-4.4	65	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026250.D
 Acq On : 08 Dec 2025 13:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.935	-2.3	67	0.01
49	Acenaphthylene	40.000	42.160	-5.4	67	0.00
50	Dimethylphthalate	40.000	41.709	-4.3	67	0.00
51	2,6-Dinitrotoluene	40.000	42.310	-5.8	65	-0.01
52 C	Acenaphthene	40.000	40.780	-2.0	66	0.00
53	3-Nitroaniline	40.000	39.474	1.3	65	-0.01
54 P	2,4-Dinitrophenol	40.000	41.617	-4.0	185	0.00
55	Dibenzofuran	40.000	42.444	-6.1	67	0.00
56 P	4-Nitrophenol	40.000	36.189	9.5	66	0.00
57	2,4-Dinitrotoluene	40.000	42.827	-7.1	68	0.00
58	Fluorene	40.000	43.069	-7.7	68	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	42.813	-7.0	68	-0.02
60	Diethylphthalate	40.000	41.924	-4.8	68	0.00
61	4-Chlorophenyl-phenylether	40.000	43.337	-8.3	67	-0.01
62	4-Nitroaniline	40.000	39.960	0.1	68	-0.01
63	Azobenzene	40.000	40.706	-1.8	67	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	65	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.207	-5.5	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.075	-5.2	68	-0.01
67	4-Bromophenyl-phenylether	40.000	42.951	-7.4	67	-0.02
68	Hexachlorobenzene	40.000	44.029	-10.1	68	-0.01
69	Atrazine	40.000	41.177	-2.9	67	0.00
70 C	Pentachlorophenol	40.000	41.222	-3.1	67	0.00
71	Phenanthrene	40.000	42.255	-5.6	69	0.00
72	Anthracene	40.000	41.916	-4.8	67	0.00
73	Carbazole	40.000	40.651	-1.6	67	0.00
74	Di-n-butylphthalate	40.000	43.077	-7.7	69	-0.02
75 C	Fluoranthene	40.000	42.176	-5.4	67	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	60	0.00
77	Benzidine	40.000	36.944	7.6	60	-0.01
78	Pyrene	40.000	42.781	-7.0	65	-0.01
79 S	Terphenyl-d14	80.000	90.743	-13.4	68	0.00
80	Butylbenzylphthalate	40.000	42.328	-5.8	64	0.00
81	Benzo(a)anthracene	40.000	41.550	-3.9	62	0.00
82	3,3'-Dichlorobenzidine	40.000	40.936	-2.3	61	-0.01
83	Chrysene	40.000	41.715	-4.3	62	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.146	-5.4	65	0.00
85 c	Di-n-octyl phthalate	40.000	42.221	-5.6	63	0.00
86 I	Perylene-d12	20.000	20.000	0.0	58	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.493	-3.7	57	-0.02
88	Benzo(b)fluoranthene	40.000	40.270	-0.7	58	0.00
89	Benzo(k)fluoranthene	40.000	41.678	-4.2	60	-0.01
90 C	Benzo(a)pyrene	40.000	41.488	-3.7	59	-0.01
91	Dibenzo(a,h)anthracene	40.000	41.557	-3.9	57	-0.01
92	Benzo(g,h,i)perylene	40.000	41.295	-3.2	56	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
Data File : BP026250.D
Acq On : 08 Dec 2025 13:44
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 08 14:04:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0