

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026256.D
 Acq On : 08 Dec 2025 17:49
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 18:09:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.506	0.547	-8.1	85	0.00
3	Pyridine	1.453	1.446	0.5	79	0.00
4	n-Nitrosodimethylamine	0.542	0.547	-0.9	81	0.00
5 S	2-Fluorophenol	1.244	1.259	-1.2	78	0.00
6	Aniline	2.061	1.930	6.4	78	0.00
7 S	Phenol-d6	1.570	1.518	3.3	77	0.00
8	2-Chlorophenol	1.414	1.414	0.0	77	0.00
9	Benzaldehyde	0.847	1.058	-24.9	80	0.00
10 C	Phenol	1.689	1.615	4.4	77	0.00
11	bis(2-Chloroethyl)ether	1.324	1.268	4.2	76	0.00
12	1,3-Dichlorobenzene	1.530	1.593	-4.1	79	0.00
13 C	1,4-Dichlorobenzene	1.541	1.604	-4.1	79	0.00
14	1,2-Dichlorobenzene	1.480	1.514	-2.3	78	0.00
15	Benzyl Alcohol	1.142	1.111	2.7	80	0.00
16	2,2'-oxybis(1-Chloropropane	1.848	1.787	3.3	80	0.00
17	2-Methylphenol	1.146	1.109	3.2	79	0.00
18	Hexachloroethane	0.560	0.599	-7.0	81	0.00
19 P	n-Nitroso-di-n-propylamine	0.968	0.929	4.0	80	0.00
20	3+4-Methylphenols	1.583	1.501	5.2	79	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
22	Acetophenone	0.507	0.522	-3.0	79	0.00
23 S	Nitrobenzene-d5	0.351	0.371	-5.7	80	0.00
24	Nitrobenzene	0.355	0.371	-4.5	79	0.00
25	Isophorone	0.689	0.719	-4.4	83	0.00
26 C	2-Nitrophenol	0.180	0.196	-8.9	80	0.00
27	2,4-Dimethylphenol	0.318	0.314	1.3	85	0.00
28	bis(2-Chloroethoxy)methane	0.433	0.439	-1.4	80	0.00
29 C	2,4-Dichlorophenol	0.324	0.344	-6.2	79	0.00
30	1,2,4-Trichlorobenzene	0.333	0.362	-8.7	79	0.00
31	Naphthalene	1.081	1.124	-4.0	79	0.00
32	Benzoic acid	0.255	0.314	-23.1	146	0.03
33	4-Chloroaniline	0.453	0.459	-1.3	81	0.00
34 C	Hexachlorobutadiene	0.217	0.243	-12.0	79	0.00
35	Caprolactam	0.114	0.116	-1.8	82	0.00
36 C	4-Chloro-3-methylphenol	0.341	0.356	-4.4	83	0.00
37	2-Methylnaphthalene	0.762	0.792	-3.9	80	0.00
38	1-Methylnaphthalene	0.744	0.786	-5.6	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	79	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.611	0.653	-6.9	81	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.411	-7.9	106	0.00
42 S	2,4,6-Tribromophenol	0.260	0.287	-10.4	84	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.445	-7.2	83	0.00
44	2,4,5-Trichlorophenol	0.462	0.494	-6.9	84	0.00
45 S	2-Fluorobiphenyl	1.376	1.480	-7.6	82	0.00
46	1,1'-Biphenyl	1.510	1.556	-3.0	81	0.00
47	2-Chloronaphthalene	1.190	1.236	-3.9	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
 Data File : BP026256.D
 Acq On : 08 Dec 2025 17:49
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 08 18:09:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.326	0.342	-4.9	86	0.00
49	Acenaphthylene	1.953	2.045	-4.7	83	0.00
50	Dimethylphthalate	1.490	1.573	-5.6	84	0.00
51	2,6-Dinitrotoluene	0.295	0.324	-9.8	84	0.00
52 C	Acenaphthene	1.142	1.168	-2.3	82	0.00
53	3-Nitroaniline	0.344	0.345	-0.3	82	-0.01
54 P	2,4-Dinitrophenol	0.157	0.207	-31.8#	291#	0.00
55	Dibenzofuran	1.776	1.867	-5.1	82	0.00
56 P	4-Nitrophenol	0.307	0.289	5.9	85	0.00
57	2,4-Dinitrotoluene	0.438	0.469	-7.1	84	0.00
58	Fluorene	1.405	1.494	-6.3	84	-0.01
59	2,3,4,6-Tetrachlorophenol	0.391	0.417	-6.6	84	-0.02
60	Diethylphthalate	1.495	1.570	-5.0	85	0.00
61	4-Chlorophenyl-phenylether	0.700	0.748	-6.9	82	0.00
62	4-Nitroaniline	0.357	0.339	5.0	80	0.00
63	Azobenzene	1.257	1.297	-3.2	85	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.142	-20.3	139	0.00
66 c	n-Nitrosodiphenylamine	0.618	0.654	-5.8	83	-0.01
67	4-Bromophenyl-phenylether	0.222	0.244	-9.9	83	-0.02
68	Hexachlorobenzene	0.258	0.287	-11.2	84	0.00
69	Atrazine	0.233	0.248	-6.4	84	-0.01
70 C	Pentachlorophenol	0.179	0.193	-7.8	86	0.00
71	Phenanthrene	1.127	1.175	-4.3	82	0.00
72	Anthracene	1.148	1.208	-5.2	82	-0.01
73	Carbazole	1.087	1.101	-1.3	81	0.00
74	Di-n-butylphthalate	1.284	1.410	-9.8	86	-0.01
75 C	Fluoranthene	1.372	1.450	-5.7	82	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
77	Benzidine	0.628	0.606	3.5	79	-0.02
78	Pyrene	1.303	1.381	-6.0	81	-0.01
79 S	Terphenyl-d14	0.979	1.086	-10.9	83	-0.01
80	Butylbenzylphthalate	0.572	0.616	-7.7	82	0.00
81	Benzo(a)anthracene	1.372	1.404	-2.3	77	0.00
82	3,3'-Dichlorobenzidine	0.499	0.514	-3.0	77	-0.02
83	Chrysene	1.299	1.363	-4.9	77	0.00
84	Bis(2-ethylhexyl)phthalate	0.860	0.926	-7.7	83	0.00
85 c	Di-n-octyl phthalate	1.517	1.660	-9.4	82	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	72	-0.02
87	Indeno(1,2,3-cd)pyrene	1.512	1.593	-5.4	72	-0.04
88	Benzo(b)fluoranthene	1.215	1.291	-6.3	76	-0.02
89	Benzo(k)fluoranthene	1.216	1.268	-4.3	75	-0.02
90 C	Benzo(a)pyrene	1.153	1.209	-4.9	75	-0.02
91	Dibenzo(a,h)anthracene	1.236	1.296	-4.9	72	-0.03
92	Benzo(g,h,i)perylene	1.232	1.308	-6.2	72	-0.07

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120825\
Data File : BP026256.D
Acq On : 08 Dec 2025 17:49
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 08 18:09:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0