

Surrogate Summary

SW-846

SDG No.: Q3776

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170839BL	PB170839BL	2-Fluorophenol	150	136	91		19	119
		Phenol-d6	150	130	87		10	130
		Nitrobenzene-d5	100	89.1	89		44	120
		2-Fluorobiphenyl	100	90.0	90		44	119
		2,4,6-Tribromophenol	150	141	94		43	140
		Terphenyl-d14	100	95.2	95		50	134
PB170839BS	PB170839BS	2-Fluorophenol	150	125	83		19	119
		Phenol-d6	150	122	81		10	130
		Nitrobenzene-d5	100	80.6	81		44	120
		2-Fluorobiphenyl	100	82.5	83		44	119
		2,4,6-Tribromophenol	150	134	89		43	140
		Terphenyl-d14	100	88.2	88		50	134
PB170839BSD	PB170839BSD	2-Fluorophenol	150	124	82		19	119
		Phenol-d6	150	120	80		10	130
		Nitrobenzene-d5	100	79.9	80		44	120
		2-Fluorobiphenyl	100	81.8	82		44	119
		2,4,6-Tribromophenol	150	137	91		43	140
		Terphenyl-d14	100	85.6	86		50	134
Q3776-01	IDW-RAD-20251203	2-Fluorophenol	150	68.6	46		19	119
		Phenol-d6	150	42.6	28		10	130
		Nitrobenzene-d5	100	87.3	87		44	120
		2-Fluorobiphenyl	100	85.7	86		44	119
		2,4,6-Tribromophenol	150	143	96		43	140
		Terphenyl-d14	100	85.7	86		50	134