

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

RunNo.: LB138120

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.1	10	101	90-110	12/01/2025
Chloride	mg/L	3	3	100	90-110	12/01/2025
Fluoride	mg/L	2	2	100	90-110	12/01/2025
Nitrite	mg/L	3	3	100	90-110	12/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	12/01/2025
Sulfate	mg/L	14.9	15	99	90-110	12/01/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	12/01/2025
Sample ID: CCV1						
Bromide	mg/L	10	10	100	90-110	12/04/2025
Chloride	mg/L	3	3	100	90-110	12/04/2025
Fluoride	mg/L	2	2	100	90-110	12/04/2025
Nitrite	mg/L	3	3	100	90-110	12/04/2025
Nitrate	mg/L	2.5	2.5	100	90-110	12/04/2025
Sulfate	mg/L	14.7	15	98	90-110	12/04/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	12/04/2025
Sample ID: CCV2						
Bromide	mg/L	9.4	10	94	90-110	12/04/2025
Chloride	mg/L	2.9	3	97	90-110	12/04/2025
Fluoride	mg/L	1.9	2	95	90-110	12/04/2025
Nitrite	mg/L	2.8	3	93	90-110	12/04/2025
Nitrate	mg/L	2.4	2.5	96	90-110	12/04/2025
Sulfate	mg/L	14.1	15	94	90-110	12/04/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	12/04/2025

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

RunNo.: LB138131

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	5	5	100	90-110	12/05/2025
Sample ID: CCV1 TKN	mg/L	5.1	5	102	90-110	12/05/2025
Sample ID: CCV2 TKN	mg/L	5	5	100	90-110	12/05/2025
Sample ID: CCV3 TKN	mg/L	5.1	5	102	90-110	12/05/2025

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

RunNo.: LB138132

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.522	0.50	104	90-110	12/05/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.499	0.50	100	90-110	12/05/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.507	0.50	101	90-110	12/05/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.498	0.50	100	90-110	12/05/2025

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

RunNo.: LB138132

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------