

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

Run No.: LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

Run No.: LB138120

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138120BSW							
Bromide	mg/L	10	9.50		95	1	90-110	12/04/2025
Chloride	mg/L	3	2.90		97	1	90-110	12/04/2025
Fluoride	mg/L	2	1.90		95	1	90-110	12/04/2025
Nitrite	mg/L	3	2.90		97	1	90-110	12/04/2025
Nitrate	mg/L	2.5	2.40		96	1	90-110	12/04/2025
Sulfate	mg/L	15	14.1		94	1	90-110	12/04/2025
Orthophosphate as P	mg/L	5	4.90		98	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

Run No.: LB138150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138150BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

Run No.: LB138131

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170831BS							
TKN	mg/L	5	5.00		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3777

Project: AEC-2025-0013

Run No.: LB138132

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170844BS							
Phosphorus, Total	mg/L	0.50	0.52		104	1	90-110	12/05/2025