

Preparation Blank Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3778

Project: AEC-2025-0013

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB138118BL							
TSS	mg/L	1	2.0000	J	1	4	12/04/2025
Sample ID: LB138120BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	12/04/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	12/04/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	12/04/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	12/04/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	12/04/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	12/04/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	12/04/2025
Sample ID: LB138150BL							
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	12/08/2025
Sample ID: PB170831BL							
TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	12/05/2025
Sample ID: PB170844BL							
Phosphorus, Total	mg/L	0.009	0.0250	J	0.005	0.05	12/05/2025