

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN112425**

|              |               |                   |                                           |
|--------------|---------------|-------------------|-------------------------------------------|
| Review By    | John Carlone  | Review On         | 11/25/2025 9:36:43 AM                     |
| Supervise By | Maresh Dadoda | Supervise On      | 11/26/2025 1:52:19 PM                     |
| SubDirectory | VN112425      | HP Acquire Method | MSVOA_N HP Processing Method 82N112425W.M |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP136378                                              |
| Initial Calibration Stds | VP136444,VP136445,VP136446,VP136447,VP136448,VP136449 |
| CCC                      |                                                       |
| Internal Standard/PEM    |                                                       |
| ICV/I.BLK                | VP136450                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                  | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|------------------------------------------|----------|--------|
| 1   | BFB         | BFB         | VN088343.D     | 24 Nov 2025 08:17 |                                          | JC\MD    | Ok     |
| 2   | VSTDICC001  | VSTDICC001  | VN088344.D     | 24 Nov 2025 12:38 |                                          | JC\MD    | Ok,M   |
| 3   | VSTDICC005  | VSTDICC005  | VN088345.D     | 24 Nov 2025 13:11 | COM.#74 Peak not detected in 1 and 5 PPB | JC\MD    | Ok,M   |
| 4   | VSTDICC020  | VSTDICC020  | VN088346.D     | 24 Nov 2025 13:32 |                                          | JC\MD    | Ok,M   |
| 5   | VSTDICCC050 | VSTDICCC050 | VN088347.D     | 24 Nov 2025 13:53 | Method failed for Comp. #74              | JC\MD    | Ok,M   |
| 6   | VSTDICC100  | VSTDICC100  | VN088348.D     | 24 Nov 2025 14:14 |                                          | JC\MD    | Ok,M   |
| 7   | VSTDICC150  | VSTDICC150  | VN088349.D     | 24 Nov 2025 14:35 |                                          | JC\MD    | Ok,M   |
| 8   | IBLK        | IBLK        | VN088350.D     | 24 Nov 2025 14:56 |                                          | JC\MD    | Ok     |
| 9   | VSTDICV050  | ICVVN112425 | VN088351.D     | 24 Nov 2025 15:18 |                                          | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN120925**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 12/10/2025 8:05:14 AM             |
| Supervise By | Maresh Dadoda | Supervise On      | 12/11/2025 11:46:02 AM            |
| SubDirectory | VN120925      | HP Acquire Method | HP Processing Method 82N112425W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP136586          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP136587,VP136588 |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment        | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|----------------|----------|----------|
| 1   | BFB          | BFB          | VN088470.D     | 09 Dec 2025 10:07 |                | JC\MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VN088471.D     | 09 Dec 2025 10:47 | pH#Lot#V12668  | JC\MD    | Ok,M     |
| 3   | VN1209MBL01  | VN1209MBL01  | VN088472.D     | 09 Dec 2025 11:21 |                | JC\MD    | Ok       |
| 4   | VN1209WBL01  | VN1209WBL01  | VN088473.D     | 09 Dec 2025 11:42 |                | JC\MD    | Ok       |
| 5   | VN1209MBS01  | VN1209MBS01  | VN088474.D     | 09 Dec 2025 12:51 |                | JC\MD    | Ok,M     |
| 6   | VN1209MBSD01 | VN1209MBSD01 | VN088475.D     | 09 Dec 2025 13:25 |                | JC\MD    | Ok,M     |
| 7   | Q3770-01     | 4207         | VN088476.D     | 09 Dec 2025 13:45 | Need 50X       | JC\MD    | Dilution |
| 8   | Q3784-03     | SB3A         | VN088477.D     | 09 Dec 2025 14:05 |                | JC\MD    | Ok,M     |
| 9   | Q3809-01     | SW-1         | VN088478.D     | 09 Dec 2025 14:26 |                | JC\MD    | Ok,M     |
| 10  | VN1209WBS01  | VN1209WBS01  | VN088479.D     | 09 Dec 2025 15:22 | Surrogate fail | JC\MD    | Not Ok   |
| 11  | Q3795-04     | 72-11966     | VN088480.D     | 09 Dec 2025 15:42 | vial A pH#5.0  | JC\MD    | Ok       |
| 12  | Q3795-08     | RBR-200059   | VN088481.D     | 09 Dec 2025 16:02 | vial A pH#5.0  | JC\MD    | Ok       |
| 13  | PB170858TB   | PB170858TB   | VN088482.D     | 09 Dec 2025 16:23 |                | JC\MD    | Ok,M     |
| 14  | VN1209WBS02  | VN1209WBS02  | VN088483.D     | 09 Dec 2025 16:43 |                | JC\MD    | Ok,M     |
| 15  | VN1209WBSD02 | VN1209WBSD02 | VN088484.D     | 09 Dec 2025 17:03 |                | JC\MD    | Not Ok   |
| 16  | VSTDCCC050   | VSTDCCC050EC | VN088485.D     | 09 Dec 2025 17:23 |                | JC\MD    | Ok,M     |

M : Manual Integration

Instrument ID: MSVOA\_W

**Daily Analysis Runlog For Sequence/QC Batch ID # VW120325**

|              |                     |                   |                                           |
|--------------|---------------------|-------------------|-------------------------------------------|
| Review By    | John Carlone        | Review On         | 12/4/2025 8:41:58 AM                      |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 12/4/2025 2:38:27 PM                      |
| SubDirectory | VW120325            | HP Acquire Method | MSVOA_W HP Processing Method 82W120325S.M |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP136481                                              |
| Initial Calibration Stds | VP136482,VP136483,VP136484,VP136485,VP136486,VP136487 |
| CCC                      |                                                       |
| Internal Standard/PEM    |                                                       |
| ICV/I.BLK                | VP136488                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | SampleID    | ClientID     | Data File Name | Date-Time         | Comment   | Operator | Status |
|-----|-------------|--------------|----------------|-------------------|-----------|----------|--------|
| 1   | BFB         | BFB          | VW032560.D     | 03 Dec 2025 09:27 |           | SY/MD    | Ok     |
| 2   | VSTDCCC050  | VSTDCCC050   | VW032561.D     | 03 Dec 2025 09:58 | Need ICAL | SY/MD    | Not Ok |
| 3   | VIBLK       | VIBLK        | VW032562.D     | 03 Dec 2025 10:45 |           | SY/MD    | Not Ok |
| 4   | VSTDICC005  | VSTDICC005   | VW032563.D     | 03 Dec 2025 11:32 |           | SY/MD    | Ok,M   |
| 5   | VSTDICC010  | VSTDICC010   | VW032564.D     | 03 Dec 2025 11:53 |           | SY/MD    | Ok,M   |
| 6   | VSTDICC020  | VSTDICC020   | VW032565.D     | 03 Dec 2025 12:15 |           | SY/MD    | Ok,M   |
| 7   | VSTDICCC050 | VSTDICCC050  | VW032566.D     | 03 Dec 2025 12:37 |           | SY/MD    | Ok,M   |
| 8   | VSTDICC100  | VSTDICC100   | VW032567.D     | 03 Dec 2025 12:58 |           | SY/MD    | Ok,M   |
| 9   | VSTDICC150  | VSTDICC150   | VW032568.D     | 03 Dec 2025 13:20 |           | SY/MD    | Ok,M   |
| 10  | VIBLK       | VIBLK        | VW032569.D     | 03 Dec 2025 13:42 |           | SY/MD    | Not Ok |
| 11  | VSTDICV050  | ICVWVW120325 | VW032570.D     | 03 Dec 2025 14:07 |           | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_W

**Daily Analysis Runlog For Sequence/QC Batch ID # VW120825**

|              |                     |                   |                                           |
|--------------|---------------------|-------------------|-------------------------------------------|
| Review By    | John Carlone        | Review On         | 12/9/2025 1:30:03 PM                      |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 12/9/2025 6:10:17 PM                      |
| SubDirectory | VW120825            | HP Acquire Method | MSVOA_W HP Processing Method 82W120325S.M |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP136572                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP136573,VP136574<br>VP136146 |

| Sr# | SampleID     | ClientID        | Data File Name | Date-Time         | Comment                                      | Operator | Status |
|-----|--------------|-----------------|----------------|-------------------|----------------------------------------------|----------|--------|
| 1   | BFB          | BFB             | VW032582.D     | 08 Dec 2025 08:11 |                                              | SY/MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050      | VW032583.D     | 08 Dec 2025 09:40 |                                              | SY/MD    | Ok,M   |
| 3   | VW1208SBL01  | VW1208SBL01     | VW032584.D     | 08 Dec 2025 10:18 |                                              | SY/MD    | Ok     |
| 4   | VW1208SBS01  | VW1208SBS01     | VW032585.D     | 08 Dec 2025 11:01 |                                              | SY/MD    | Ok,M   |
| 5   | VW1208SBS02  | VW1208SBS02     | VW032586.D     | 08 Dec 2025 11:22 | Surrogate fail;Not Required                  | SY/MD    | Not Ok |
| 6   | Q3784-04     | SB4A            | VW032587.D     | 08 Dec 2025 11:56 | vial B                                       | SY/MD    | Ok     |
| 7   | Q3801-03     | SOIL-1-VOC      | VW032588.D     | 08 Dec 2025 12:18 | vial A                                       | SY/MD    | Ok     |
| 8   | VW1208SBSD01 | VW1208SBSD01    | VW032589.D     | 08 Dec 2025 12:39 |                                              | SY/MD    | Ok,M   |
| 9   | Q3801-08     | SOIL#2-VOC      | VW032590.D     | 08 Dec 2025 13:01 | vial A                                       | SY/MD    | Ok     |
| 10  | Q3806-06     | TR-05-12-5-2025 | VW032591.D     | 08 Dec 2025 13:22 | vial A                                       | SY/MD    | Ok     |
| 11  | Q3790-03     | 304 FURMAN SOIL | VW032592.D     | 08 Dec 2025 13:44 | Internal standard fail;Surrogate fail,vial A | SY/MD    | Ok     |
| 12  | Q3795-03     | 72-11966-VOC    | VW032593.D     | 08 Dec 2025 14:05 | Surrogate fail,vial A                        | SY/MD    | Ok     |
| 13  | Q3795-07     | RBR-200059-VOC  | VW032594.D     | 08 Dec 2025 14:27 | Internal Standard Fail,vial A                | SY/MD    | ReRun  |
| 14  | Q3806-03     | TR-01-12-5-2025 | VW032595.D     | 08 Dec 2025 14:49 | vial A                                       | SY/MD    | Ok     |
| 15  | VIBLK        | VIBLK           | VW032596.D     | 08 Dec 2025 15:10 |                                              | SY/MD    | Not Ok |
| 16  | Q3804-02     | 120325-A        | VW032597.D     | 08 Dec 2025 15:32 | Surrogate fail,vial A                        | SY/MD    | ReRun  |
| 17  | VIBLK        | VIBLK           | VW032598.D     | 08 Dec 2025 15:53 |                                              | SY/MD    | Ok     |

Instrument ID: MSVOA\_W

**Daily Analysis Runlog For Sequence/QC Batch ID # VW120825**

| Review By                               | John Carlone        | Review On         | 12/9/2025 1:30:03 PM                      |
|-----------------------------------------|---------------------|-------------------|-------------------------------------------|
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 12/9/2025 6:10:17 PM                      |
| SubDirectory                            | VW120825            | HP Acquire Method | MSVOA_W HP Processing Method 82W120325S.M |
| STD. NAME                               | STD REF.#           |                   |                                           |
| Tune/Reschk<br>Initial Calibration Stds | VP136572            |                   |                                           |
| CCC                                     | VP136573,VP136574   |                   |                                           |
| Internal Standard/PEM                   | VP136146            |                   |                                           |
| ICV/I.BLK                               |                     |                   |                                           |
| Surrogate Standard                      |                     |                   |                                           |
| MS/MSD Standard                         |                     |                   |                                           |
| LCS Standard                            |                     |                   |                                           |

|    |            |              |            |                   |                               |       |      |
|----|------------|--------------|------------|-------------------|-------------------------------|-------|------|
| 18 | Q3809-02   | SW-2         | VW032599.D | 08 Dec 2025 16:15 | Internal Standard Fail,vial A | SY/MD | Ok   |
| 19 | Q3809-03   | B-1          | VW032600.D | 08 Dec 2025 16:36 | vial A                        | SY/MD | Ok,M |
| 20 | Q3809-04   | SW-3         | VW032601.D | 08 Dec 2025 16:58 | vial A                        | SY/MD | Ok,M |
| 21 | Q3809-05   | SW-4         | VW032602.D | 08 Dec 2025 17:19 | vial A                        | SY/MD | Ok   |
| 22 | VSTDCCC050 | VSTDCCC050EC | VW032603.D | 08 Dec 2025 17:41 |                               | SY/MD | Ok,M |

M : Manual Integration