

Surrogate Summary

SDG No.: **Q3784**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3784-03	SB3A	1,2-Dichloroethane-d4	50	59.6	119		70 (63)	130 (155)
		Dibromofluoromethane	50	50.5	101		70 (70)	130 (134)
		Toluene-d8	50	56.5	113		70 (74)	130 (123)
		4-Bromofluorobenzene	50	67.4	135	*	70 (52)	130 (138)
Q3784-04	SB4A	1,2-Dichloroethane-d4	50	48.4	97		70 (63)	130 (155)
		Dibromofluoromethane	50	45.3	91		70 (70)	130 (134)
		Toluene-d8	50	40.2	80		70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.4	79		70 (52)	130 (138)
VN1209MBL01	VN1209MBL01	1,2-Dichloroethane-d4	50	67.3	135	*	70 (63)	130 (155)
		Dibromofluoromethane	50	54.3	109		70 (70)	130 (134)
		Toluene-d8	50	51.2	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	54.0	108		70 (52)	130 (138)
VN1209MBS01	VN1209MBS01	1,2-Dichloroethane-d4	50	58.3	117		70 (63)	130 (155)
		Dibromofluoromethane	50	55.3	111		70 (70)	130 (134)
		Toluene-d8	50	53.1	106		70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.9	106		70 (52)	130 (138)
VW1208SBL01	VW1208SBL01	1,2-Dichloroethane-d4	50	54.2	108		70 (63)	130 (155)
		Dibromofluoromethane	50	50.7	101		70 (70)	130 (134)
		Toluene-d8	50	49.3	99		70 (74)	130 (123)
		4-Bromofluorobenzene	50	45.1	90		70 (52)	130 (138)
VW1208SBS01	VW1208SBS01	1,2-Dichloroethane-d4	50	50.1	100		70 (63)	130 (155)
		Dibromofluoromethane	50	48.5	97		70 (70)	130 (134)
		Toluene-d8	50	47.7	95		70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.3	95		70 (52)	130 (138)
VW1208SBSD0	VW1208SBSD01	1,2-Dichloroethane-d4	50	55.6	111		70 (63)	130 (155)
		Dibromofluoromethane	50	54.4	109		70 (70)	130 (134)
		Toluene-d8	50	55.7	111		70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.4	107		70 (52)	130 (138)

