

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3784</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VW032585.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1208SBS01	Dichlorodifluoromethane	20	22.5	ug/Kg	113			40 (64)	160 (136)	
	Chloromethane	20	18.8	ug/Kg	94			40 (73)	160 (129)	
	Vinyl chloride	20	20.0	ug/Kg	100			70 (73)	130 (133)	
	Bromomethane	20	19.6	ug/Kg	98			40 (77)	160 (137)	
	Chloroethane	20	19.5	ug/Kg	98			40 (69)	160 (130)	
	Trichlorofluoromethane	20	20.8	ug/Kg	104			40 (69)	160 (134)	
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/Kg	103			70 (81)	130 (123)	
	Tert butyl alcohol	100	89.1	ug/Kg	89			70 (24)	130 (175)	
	1,1-Dichloroethene	20	19.6	ug/Kg	98			70 (79)	130 (121)	
	Acetone	100	110	ug/Kg	110			40 (60)	160 (174)	
	Carbon disulfide	20	19.9	ug/Kg	100			40 (73)	160 (125)	
	Methyl tert-butyl Ether	20	20.8	ug/Kg	104			70 (77)	130 (129)	
	Methyl Acetate	20	20.9	ug/Kg	104			70 (51)	130 (140)	
	Methylene Chloride	20	19.3	ug/Kg	97			70 (54)	130 (158)	
	trans-1,2-Dichloroethene	20	19.5	ug/Kg	98			70 (80)	130 (123)	
	1,1-Dichloroethane	20	20.1	ug/Kg	101			70 (82)	130 (123)	
	Cyclohexane	20	20.1	ug/Kg	101			70 (76)	130 (122)	
	2-Butanone	100	100	ug/Kg	100			40 (69)	160 (131)	
	Carbon Tetrachloride	20	19.6	ug/Kg	98			70 (76)	130 (129)	
	cis-1,2-Dichloroethene	20	19.1	ug/Kg	96			70 (82)	130 (123)	
	Bromochloromethane	20	20.6	ug/Kg	103			70 (80)	130 (127)	
	Chloroform	20	19.9	ug/Kg	100			70 (82)	130 (125)	
	1,1,1-Trichloroethane	20	20.0	ug/Kg	100			70 (80)	130 (126)	
	Methylcyclohexane	20	19.3	ug/Kg	97			70 (77)	130 (123)	
	Benzene	20	19.1	ug/Kg	96			70 (84)	130 (121)	
	1,2-Dichloroethane	20	20.0	ug/Kg	100			70 (81)	130 (126)	
	Trichloroethene	20	19.4	ug/Kg	97			70 (83)	130 (122)	
	1,2-Dichloropropane	20	19.9	ug/Kg	100			70 (83)	130 (122)	
	Bromodichloromethane	20	19.7	ug/Kg	99			70 (82)	130 (123)	
	4-Methyl-2-Pentanone	100	110	ug/Kg	110			40 (70)	160 (135)	
	Toluene	20	19.7	ug/Kg	99			70 (83)	130 (122)	
	t-1,3-Dichloropropene	20	19.2	ug/Kg	96			70 (78)	130 (124)	
	cis-1,3-Dichloropropene	20	19.3	ug/Kg	97			70 (81)	130 (122)	
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103			70 (82)	130 (125)	
	2-Hexanone	100	110	ug/Kg	110			40 (66)	160 (138)	
	Dibromochloromethane	20	19.4	ug/Kg	97			70 (79)	130 (125)	
	1,2-Dibromoethane	20	20.6	ug/Kg	103			70 (80)	130 (125)	
	Tetrachloroethene	20	18.1	ug/Kg	91			70 (83)	130 (125)	
	Chlorobenzene	20	17.9	ug/Kg	90			70 (84)	130 (122)	
	Ethyl Benzene	20	18.2	ug/Kg	91			70 (82)	130 (124)	
	m/p-Xylenes	40	37.4	ug/Kg	94			70 (83)	130 (124)	
	o-Xylene	20	18.0	ug/Kg	90			70 (83)	130 (123)	
	Styrene	20	19.3	ug/Kg	97			70 (82)	130 (124)	
	Bromoform	20	19.1	ug/Kg	96			70 (75)	130 (127)	
	Isopropylbenzene	20	18.7	ug/Kg	94			70 (82)	130 (124)	
	1,1,2,2-Tetrachloroethane	20	20.5	ug/Kg	103			70 (77)	130 (127)	
	1,3-Dichlorobenzene	20	19.3	ug/Kg	97			70 (83)	130 (122)	
	1,4-Dichlorobenzene	20	18.9	ug/Kg	95			70 (84)	130 (121)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3784</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VW032585.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1208SBS01	1,2-Dichlorobenzene	20	19.7	ug/Kg	99			70 (83)	130 (124)	
	1,2-Dibromo-3-Chloropropane	20	19.8	ug/Kg	99			40 (66)	160 (134)	
	1,2,4-Trichlorobenzene	20	17.3	ug/Kg	86			70 (78)	130 (127)	
	1,2,3-Trichlorobenzene	20	17.9	ug/Kg	90			70 (70)	130 (137)	