

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3784
 Lab File ID: VN088471.D Date Analyzed: 12/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	234210	8.20	421047	9.08	382607	11.84
UPPER LIMIT	468420	8.7	842094	9.577	765214	12.341
LOWER LIMIT	117105	7.7	210524	8.577	191304	11.341
EPA SAMPLE NO.						
SB3A	229100	8.21	515203	9.08	538236	11.84
VN1209MBL01	179224	8.21	426527	9.08	423766	11.84
VN1209MBS01	227429	8.20	428540	9.08	382331	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.