

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3784
 Lab File ID: VN088471.D Date Analyzed: 12/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	183020	13.765				
UPPER LIMIT	366040	14.265				
LOWER LIMIT	91510	13.265				
EPA SAMPLE NO.						
SB3A	266826	13.77				
VN1209MBL01	191763	13.77				
VN1209MBS01	181539	13.76				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.