



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB138119

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.386	0.4	96	90-110	12/04/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.397	0.4	99	90-110	12/04/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.386	0.4	96	90-110	12/04/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB138124

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
COD		mg/L	50.173	50	100	95-105	09/08/2025
Sample ID:	CCV1						
COD		mg/L	48.150	50	96	95-105	12/05/2025
Sample ID:	CCV2						
COD		mg/L	48.150	50	96	95-105	12/05/2025
Sample ID:	CCV3						
COD		mg/L	49.161	50	98	95-105	12/05/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB138125

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/05/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB138153

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.098	0.099	99	90-110	12/08/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	12/08/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	12/08/2025
Sample ID: Cyanide	CCV3	mg/L	0.26	0.25	104	90-110	12/08/2025