

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138119

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138119BS							
Residual Chlorine	mg/L	0.4	0.42		105	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138124

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138124BS							
COD	mg/L	50	50.2		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138133

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138133BS							
BOD5	mg/L	198	190		96	1	84.6-115.4	12/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138150BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138125

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170830BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3785
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB138153

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170853BS							
Cyanide	mg/L	0.1	0.097		97	1	85-115	12/08/2025