



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.84	6.0	97	65 - 135	P	12/09/2025	12:28	LB138174
	Chromium	10.4	10.0	104	65 - 135	P	12/09/2025	12:28	LB138174
	Copper	21.5	20.0	107	65 - 135	P	12/09/2025	12:28	LB138174
	Lead	12.0	12.0	100	65 - 135	P	12/09/2025	12:28	LB138174
	Nickel	39.7	40.0	99	65 - 135	P	12/09/2025	12:28	LB138174
	Silver	10.0	10.0	100	65 - 135	P	12/09/2025	12:28	LB138174
	Zinc	41.9	40.0	105	65 - 135	P	12/09/2025	12:28	LB138174
CRI01	Cadmium	5.86	6.0	98	65 - 135	P	12/11/2025	11:50	LB138198
	Chromium	10.3	10.0	103	65 - 135	P	12/11/2025	11:50	LB138198
	Copper	21.8	20.0	109	65 - 135	P	12/11/2025	11:50	LB138198
	Lead	13.7	12.0	114	65 - 135	P	12/11/2025	11:50	LB138198
	Nickel	41.1	40.0	103	65 - 135	P	12/11/2025	11:50	LB138198
	Silver	11.3	10.0	113	65 - 135	P	12/11/2025	11:50	LB138198
	Zinc	43.9	40.0	110	65 - 135	P	12/11/2025	11:50	LB138198