



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	12:24	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	12:24	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	12:24	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	12:24	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	12:24	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	12:24	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	12:24	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	12:53	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	12:53	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	12:53	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	12:53	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	12:53	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	12:53	LB138174
CCB02	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	13:45	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	13:45	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	13:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	13:45	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	13:45	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	13:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	13:45	LB138174
CCB03	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	15:13	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	15:13	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	15:13	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	15:13	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	15:13	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	15:13	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	15:13	LB138174
CCB04	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	16:04	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	16:04	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	16:04	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	16:04	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	16:04	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	16:04	LB138174
CCB05	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	16:54	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	16:54	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	16:54	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	16:54	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	16:54	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	16:54	LB138174
CCB06	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	17:45	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	17:45	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	17:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	17:45	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	17:45	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	17:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	17:45	LB138174
CCB07	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	18:36	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	18:36	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	18:36	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	18:36	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	18:36	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	18:36	LB138174
CCB08	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	19:34	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	19:34	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	19:34	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	19:34	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	19:34	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	19:34	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	19:34	LB138174
CCB09	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	20:41	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	20:41	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	20:41	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	20:41	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	20:41	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	20:41	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	20:41	LB138174
CCB10	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	21:31	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	21:31	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	21:31	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	21:31	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	21:31	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	21:31	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	21:31	LB138174
CCB11	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	22:22	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	22:22	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	22:22	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	22:22	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	22:22	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	22:22	LB138174
CCB12	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	23:25	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	23:25	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	23:25	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	23:25	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	23:25	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	23:25	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	23:25	LB138174
CCB13	Cadmium	0.50	+/-3	U	6.00	P	12/09/2025	23:45	LB138174
	Chromium	2.12	+/-5	U	10.0	P	12/09/2025	23:45	LB138174
	Copper	4.60	+/-10	U	20.0	P	12/09/2025	23:45	LB138174
	Lead	2.30	+/-6	U	12.0	P	12/09/2025	23:45	LB138174
	Nickel	3.06	+/-20	U	40.0	P	12/09/2025	23:45	LB138174
	Silver	1.62	+/-5	U	10.0	P	12/09/2025	23:45	LB138174
	Zinc	16.7	+/-20	U	40.0	P	12/09/2025	23:45	LB138174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	11:46	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	11:46	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	11:46	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	11:46	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	11:46	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	11:46	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	12:18	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	12:18	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	12:18	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	12:18	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	12:18	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
CCB02	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	13:08	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	13:08	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	13:08	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	13:08	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	13:08	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
CCB03	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	13:59	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	13:59	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	13:59	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	13:59	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	13:59	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
CCB04	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	15:04	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	15:04	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	15:04	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	15:04	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	15:04	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
CCB05	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	15:54	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	15:54	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	15:54	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	15:54	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	15:54	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
CCB06	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	16:44	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	16:44	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	16:44	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	16:44	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3785

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	16:44	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
CCB07	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	17:57	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	17:57	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	17:57	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	17:57	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	17:57	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
CCB08	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	18:52	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	18:52	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	18:52	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	18:52	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	18:52	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
CCB09	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	19:45	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	19:45	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	19:45	LB138198
	Lead	3.43	+/-6	J	12.0	P	12/11/2025	19:45	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	19:45	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
CCB10	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	20:37	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	20:37	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	20:37	LB138198
	Lead	2.34	+/-6	J	12.0	P	12/11/2025	20:37	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	20:37	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
CCB11	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	20:59	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	20:59	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	20:59	LB138198
	Lead	2.32	+/-6	J	12.0	P	12/11/2025	20:59	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	20:59	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	20:59	LB138198