

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3787

Analytical Method: SW8270-Modified

Client: JACOBS Engineering Group, Inc.

DataFile: BN038338.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3787-02MS		Client Sample ID: MW-15B-42.5-120425MS									
1,4-Dioxane	0.45	0	0.20	ug/L	44				20 (10)	160 (175)	