

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3787

Client ID : SSTDCCC0.4

Date Analyzed: 12/11/2025

Lab File ID: BN038330.D

Time Analyzed: 23:59

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5325	7.724	14404	10.52	7509	14.39
UPPER LIMIT	10650	8.224	28808	11.019	15018	14.887
LOWER LIMIT	2662.5	7.224	7202	10.019	3754.5	13.887
EPA SAMPLE NO.						
01 MW-15B-42.5-120425	5971	7.72	15803	10.52	7965	14.39
02 MW-15B-42.5-120425MS	5312	7.72	13867	10.52	6982	14.39
03 MW-15B-42.5-120425MSD	5233	7.72	13599	10.52	6783	14.39
04 OWBR-05-135-120525MS	5363	7.72	14242	10.52	7200	14.39
05 OWBR-05-135-120525MSD	5563	7.72	14775	10.52	7579	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.