

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3787

Client ID: SSTDCCC0.4

Date Analyzed: 12/11/2025

Lab File ID: BN038330.D

Time Analyzed: 23:59

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	13836	17.148	10979	21.339	11587	23.63
UPPER LIMIT	27672	17.648	21958	21.839	23174	24.13
LOWER LIMIT	6918	16.648	5489.5	20.839	5793.5	23.13
EPA SAMPLE NO.						
01 MW-15B-42.5-120425	14214	17.15	10505	21.34	10318	23.63
02 MW-15B-42.5-120425MS	12326	17.15	9977	21.34	10317	23.63
03 MW-15B-42.5-120425MSD	11975	17.15	9813	21.34	10087	23.63
04 OWBR-05-135-120525MS	12520	17.15	10004	21.34	10310	23.63
05 OWBR-05-135-120525MSD	13283	17.15	10457	21.34	10825	23.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.