

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3787

Client ID : SSTDCCC0.4

Date Analyzed: 12/11/2025

Lab File ID: BN038311.D

Time Analyzed: 11:13

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6458	7.724	17315	10.53	8838	14.40
UPPER LIMIT	12916	8.224	34630	11.03	17676	14.898
LOWER LIMIT	3229	7.224	8657.5	10.03	4419	13.898
EPA SAMPLE NO.						
01 OW-03B-51.5-120425	5614	7.72	15321	10.53	7828	14.40
02 OW-03B-51.5-120425-FD	5356	7.72	14557	10.53	7569	14.40
03 OW-08B-72.5-120425	4848	7.72	13267	10.53	7308	14.40
04 OW-08B-72.5-120425-FD	4745	7.72	13148	10.53	7382	14.40
05 OW-02B-21.2-120425	5836	7.72	16262	10.52	8730	14.39
06 PB170850BS	6134	7.72	16248	10.52	7966	14.39
07 PB170850BL	6927	7.72	17922	10.53	8615	14.40
08 OWBR-02-170-120425	5142	7.72	13961	10.53	7300	14.40
09 OWBR-02-170-120425-FD	5574	7.72	15065	10.53	7765	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.