

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3787</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>12/09/2025 13:58</u>
Lab File ID:	<u>VV039501.D</u>	Init. Calib. Date(s):	<u>12/03/2025 12/03/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:55 12:18</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRFCAL	MIN RRF	%D	MAX%D
Vinyl chloride	1.082	0.964	0.01	-10.9	30
Vinyl Chloride-d3	0.491	0.543	0.01	10.5	30
1,2-Dichloroethane-d4	0.333	0.310	0.01	-7	25
1,2-Dichloropropane-d6	0.422	0.429	0.1	1.6	20
Toluene-d8	0.712	0.699	0.2	-1.7	20
1,1,2,2-Tetrachloroethane-d2	0.238	0.243	0.01	2.1	25

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.