

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3787
 Lab File ID: VV039486.D Date Analyzed: 12/08/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 (DFB) AREA #	RT #	IS2 (CBZ) AREA #	RT #	IS3 (DCB) AREA #	RT #
12 HOUR STD	6968	5.57	7021	8.79	3931	11.18
UPPER LIMIT	13936	5.736	14042	8.955	7862	11.345
LOWER LIMIT	3484	5.396	3510.5	8.615	1965.5	11.005
EPA SAMPLE NO.						
TB01-120425	6543	5.57	6120	8.79	3130	11.20
VBLK237	7024	5.57	6420	8.79	3401	11.20

IS1 (DFB) = 1,4-Difluorobenzene
 IS2 (CBZ) = Chlorobenzene-d5
 IS3 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.