

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048745.D
 Acq On : 08 Dec 2025 11:44
 Operator : JC/MD
 Sample : VX1208WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1208WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/09/2025
 Supervised By : Mahesh Dadoda 12/09/2025

Quant Time: Dec 09 04:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	178446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.738	114	298973	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	311271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.000	152	157846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.928	65	135832	56.713	ug/l	-0.01
Spiked Amount 50.000	Range 78 - 117		Recovery = 113.420%			
35) Dibromofluoromethane	5.367	113	103655	50.354	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.700%			
50) Toluene-d8	8.628	98	364834	50.560	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.120%			
62) 4-Bromofluorobenzene	11.061	95	146293	58.794	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 117.580%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.172	85	41301	22.716	ug/l	100
3) Chloromethane	1.300	50	44715	19.489	ug/l	99
4) Vinyl Chloride	1.380	62	48480	19.841	ug/l	99
5) Bromomethane	1.617	94	33887	22.441	ug/l	96
6) Chloroethane	1.697	64	30088	19.785	ug/l	96
7) Trichlorofluoromethane	1.898	101	71250	20.345	ug/l	99
8) Diethyl Ether	2.130	74	26662	19.074	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.337	101	39901	19.653	ug/l	97
10) Methyl Iodide	2.453	142	50587	17.502	ug/l	97
11) Tert butyl alcohol	2.928	59	39487	102.117	ug/l	99
12) 1,1-Dichloroethene	2.325	96	39681	19.150	ug/l	94
13) Acrolein	2.233	56	38865	127.548	ug/l	98
14) Allyl chloride	2.666	41	73696	21.184	ug/l	96
15) Acrylonitrile	3.050	53	121301	100.674	ug/l	98
16) Acetone	2.367	43	99898	108.954	ug/l	91
17) Carbon Disulfide	2.514	76	112473	18.691	ug/l	99
18) Methyl Acetate	2.690	43	55490	21.510	ug/l	96
19) Methyl tert-butyl Ether	3.099	73	137477	20.386	ug/l	100
20) Methylene Chloride	2.788	84	46445	20.266	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	38834	18.362	ug/l	97
22) Diisopropyl ether	3.745	45	138676	21.329	ug/l	99
23) Vinyl Acetate	3.709	43	614319	112.447	ug/l	95
24) 1,1-Dichloroethane	3.599	63	75325	20.775	ug/l	99
25) 2-Butanone	4.532	43	153848	113.689	ug/l #	89
26) 2,2-Dichloropropane	4.452	77	65032	20.271	ug/l	97
27) cis-1,2-Dichloroethene	4.471	96	47580	19.937	ug/l	94
28) Bromochloromethane	4.879	49	38829	21.849	ug/l	84
29) Tetrahydrofuran	4.971	42	103797	100.452	ug/l	95
30) Chloroform	5.068	83	78114	20.107	ug/l	93
31) Cyclohexane	5.452	56	60315	18.632	ug/l	99
32) 1,1,1-Trichloroethane	5.361	97	69881	20.075	ug/l	96
36) 1,1-Dichloropropene	5.672	75	50056	17.754	ug/l	99
37) Ethyl Acetate	4.690	43	72073	21.393	ug/l	96
38) Carbon Tetrachloride	5.659	117	61250	19.179	ug/l	93
39) Methylcyclohexane	7.360	83	57003	17.098	ug/l	97
40) Benzene	6.019	78	157310	18.821	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
 Data File : VX048745.D
 Acq On : 08 Dec 2025 11:44
 Operator : JC/MD
 Sample : VX1208WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1208WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 12/09/2025
 Supervised By :Mahesh Dadoda 12/09/2025

Quant Time: Dec 09 04:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.891	41	31647	20.646	ug/l	94
42) 1,2-Dichloroethane	6.068	62	63162	21.351	ug/l	93
43) Isopropyl Acetate	6.312	43	105648	20.710	ug/l	94
44) Trichloroethene	7.104	130	38336	17.578	ug/l	94
45) 1,2-Dichloropropane	7.409	63	38148	18.261	ug/l	99
46) Dibromomethane	7.562	93	29948	18.904	ug/l	95
47) Bromodichloromethane	7.799	83	62155	18.818	ug/l	97
48) Methyl methacrylate	7.671	41	49411	19.255	ug/l	94
49) 1,4-Dioxane	7.641	88	11933	338.111	ug/l	95
51) 4-Methyl-2-Pentanone	8.549	43	333290	105.257	ug/l	96
52) Toluene	8.702	92	98537	19.551	ug/l	99
53) t-1,3-Dichloropropene	8.958	75	62916	20.022	ug/l	98
54) cis-1,3-Dichloropropene	8.348	75	65243	19.320	ug/l	94
55) 1,1,2-Trichloroethane	9.134	97	44019	21.707	ug/l	99
56) Ethyl methacrylate	9.098	69	67147	21.154	ug/l	97
57) 1,3-Dichloropropane	9.287	76	69273	20.543	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.220	63	152119	88.668	ug/l	97
59) 2-Hexanone	9.409	43	232393	104.551	ug/l	98
60) Dibromochloromethane	9.500	129	53723	21.424	ug/l	100
61) 1,2-Dibromoethane	9.592	107	46114	21.010	ug/l	99
64) Tetrachloroethene	9.256	164	37431	16.498	ug/l	97
65) Chlorobenzene	10.061	112	118471	17.388	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.146	131	42596	17.479	ug/l	99
67) Ethyl Benzene	10.177	91	198185	17.826	ug/l	100
68) m/p-Xylenes	10.281	106	155889	35.995	ug/l	96
69) o-Xylene	10.622	106	73460	18.102	ug/l	97
70) Styrene	10.634	104	126903	18.215	ug/l	99
71) Bromoform	10.780	173	37658	17.309	ug/l #	99
73) Isopropylbenzene	10.945	105	188979	17.766	ug/l	99
74) N-amyl acetate	10.823	43	81515	17.637	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.195	83	63497	17.692	ug/l	99
76) 1,2,3-Trichloropropane	11.219	75	49341m	16.783	ug/l	
77) Bromobenzene	11.183	156	50590	17.819	ug/l	93
78) n-propylbenzene	11.286	91	218787	17.759	ug/l	98
79) 2-Chlorotoluene	11.347	91	131768	17.540	ug/l	97
80) 1,3,5-Trimethylbenzene	11.433	105	154012	17.615	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	20596	16.641	ug/l	90
82) 4-Chlorotoluene	11.433	91	156772	17.883	ug/l	98
83) tert-Butylbenzene	11.695	119	158909	17.491	ug/l	98
84) 1,2,4-Trimethylbenzene	11.732	105	158196	17.937	ug/l	99
85) sec-Butylbenzene	11.872	105	190539	17.306	ug/l	98
86) p-Isopropyltoluene	11.988	119	164163	17.443	ug/l	100
87) 1,3-Dichlorobenzene	11.951	146	90505	17.292	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	92718	17.143	ug/l	99
89) n-Butylbenzene	12.311	91	144814	16.434	ug/l	99
90) Hexachloroethane	12.518	117	30759	16.013	ug/l	97
91) 1,2-Dichlorobenzene	12.317	146	85738	16.897	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.920	75	13707	15.692	ug/l	95
93) 1,2,4-Trichlorobenzene	13.567	180	55758	16.638	ug/l	99
94) Hexachlorobutadiene	13.707	225	21763	15.700	ug/l	99
95) Naphthalene	13.756	128	177703	20.256	ug/l	100
96) 1,2,3-Trichlorobenzene	13.938	180	52737	17.064	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120825\
Data File : VX048745.D
Acq On : 08 Dec 2025 11:44
Operator : JC/MD
Sample : VX1208WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1208WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/09/2025
Supervised By :Mahesh Dadoda 12/09/2025

Quant Time: Dec 09 04:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

